



12 Angelier street, Bardene,  
Boksburg, 1459  
P.O.BOX 328, ISANDO, 1600  
TEL: +27 11 918 6994  
FAX: +27 11 894 1037

VAT REG #: 421 021 2249  
CO REG #: 2004/007607/07

## CALIBRATION CERTIFICATE

					<b>CERTIFICATE NO:</b>	
					<b>1353</b>	
					Doc. No: EA_5656	
Customer:	Medupi Power				Job No:	5656
Contact Name:	Wessel v Wyk	Phone:	0726119371	Fax:		
Model:	Foedisch MSA 12	Serial No:				
Auxiliary Equipment:	Unit 6					

CHAN.		PRIOR	FINAL	TEST GAS	TEST GAS VALUE	GAS CERTIFICATE NUMBER
1	ZERO	0.00 ppm	0.00 ppm	N2	99.999%	134494
	SPAN	432 ppm	450 ppm	CO	450.2ppm	2303633
2	ZERO	0.2 %	0.00 %	N2	99.999%	134494
	SPAN	21.2 %	20.95 %	O2	20.9%	Ambient Air
3	ZERO	0.00 ppm	0.00 ppm	N2	99.999%	134494
	SPAN	1100 ppm	1089 ppm	SO2	1089.1 ppm	2268715
4	ZERO	0.00 ppm	0.00 ppm	N2	99.999%	134494
	SPAN	691 ppm	690 ppm	NO	690.0 ppm	2303663

**Report : Diagnostics OK**

Date: 11 March 2015

Cal Due Date: June 2015

Service Engineer: M Langner

