



12 Angelier street, Bardene,
Boksburg, 1459
P.O.BOX 328, ISANDO, 1600
TEL: +27 11 918 6994
FAX: +27 11 894 1037

VAT REG #: 421 021 2249
CO REG #: 2004/007607/07

CALIBRATION CERTIFICATE

					CERTIFICATE NO:	
					1548	
					Doc. No: EA_5656	
Customer:	Medupi Power			Job No:	6018	
Contact Name:	Wessel v Wyk	Phone:	0726119371	Fax:		
Model:	Foedisch MGA - 23		Serial No:			
Auxiliary Equipment:	Unit 6 – MGA-23					

CHAN.		PRIOR	FINAL	TEST GAS	TEST GAS VALUE	GAS CERTIFICATE NUMBER
1	ZERO	0.18%	0.00%	Air	0.00	Air
	SPAN	721.33%	738.13ppm	CO	738.2ppm	2343712
2	ZERO	-0.16%	0.00%	N2	99.999%	2343712
	SPAN	21.18 %	20.95%	O2	20.95%	Ambient Air
3	ZERO	0.11ppm	0.00ppm	Air	0.00	Air
	SPAN	1177.6ppm	1300ppm	SO2	1300 ppm	2303623
4	ZERO	0.15ppm	0.00ppm	Air	0.00	Air
	SPAN	693.5ppm	703.96ppm	NO	704.6ppm	2343712

Report : Diagnostics OK , all calibrations done according to cylinder values.

Date: 11 September 2015

Cal Due Date: Dec 2015

Service Engineer: M Langner / J Faber

